

---

---

Date: 17 June 2020

**Clarification Note # 2**

**Invitation to Bid (ITB) No. UKR/02/2020 - Supply and Delivery of Chemical Control Kits,  
Chemical Protection and Laboratory Equipment**

---

---

In accordance with issued ITB No. UKR/02/2020 - Supply and Delivery of Chemical Control Kits, Chemical Protection and Laboratory Equipment, the OSCE would like to provide the following clarifications to the ITB Documents:

**Question 1:** After detailed studying of the tender documentation we have a question regarding Lot 2, Item IV: Filter 3M 60926: the mentioned filter does not correspond to the requirements of safety class stated below (Filter Class: A1 Hg P3 R or higher). In particular, the 3M 60926 filter has no protection against Hg. Therefore, does the customer need an equivalent filter with protection class A1 Hg P3 R or higher or the named filter 3M 60926 with lower protection is preferable?

**Answer 1:** Description of Lot II / Item IV requires the filter with class of protection minimum A1 Hg P3 R. Therefore, instead of model 3M 60926, the **filter 3M 6096A** may be used as reference: <https://multimedia.3m.com/mws/media/536278O/6096a-cartridge-filter-pis.pdf>

Sincerely,

  
Emina Sibic  
Chief of Fund Administration,  
OSCE Project Coordinator in Ukraine

